



AI Engineering Services Limited

Date: 22 July 2023

CORRIGENDUM-1

TENDER DOCUMENT FOR THE SUPPLY OF FLUORESCENT MAGNETIC PARTICLE INSPECTION (FMPI) MACHINE THROUGH BUYBACK POLICY

LAST DATE OF SUBMISSION OF BIDS	08 Aug 2023 10:30 HRS
DATE OF OPENING OF TECHNICAL BIDS	08 Aug 2023 11:30 HRS

TECHNICAL SPECIFICATIONS, Annexure B page 14 following are being included

1. The least count of FMPI machine should be 20 AMP.
2. The Current Carrying capacity should be Minimum 500 AMP (± 50 AMP) and Maximum 6000 AMP (± 300 AMP)
3. 4: FMPI Machine Current Variation should not be more than 5%.
4. 5: Bidder should have supplier approval for chemicals with QPL listing to be used on the FMPI machine.

PRICE BID Annexure B page 15 following are being included

Supply of MPFI Machine - Delivery & Commissioning address should be at New engineering Complex Gate No. 3, AIEngineering Services Ltd, Vile Parle East, near Bamanwada, Mumbai, 400099.